



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/621,490	07/17/2003	Jacqueline Coral Kent	07-2063-A	2623
20306 7590 09/11/2008 MCDONNELL BOEHNEN HULBERT & BERGHOFF LLP 300 S. WACKER DRIVE 32ND FLOOR CHICAGO, IL 60606			EXAMINER WITCZAK, CATHERINE	
			ART UNIT 3767	PAPER NUMBER
			MAIL DATE 09/11/2008	DELIVERY MODE PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary

Application No.

10/621,490

Applicant(s)

KENT ET AL.

Examiner

CATHERINE N. WITCZAK

Art Unit

3767

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --
Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 03 June 2008.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1,2,4-10 and 12-18 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1,2,4-10 and 12-18 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☒ Information Disclosure Statement(s) (PTO/S5108)
Paper No(s)/Mail Date 7/22/2008
- 4) ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date _____
- 5) ☐ Notice of Informal Patent Application
- 6) ☐ Other: _____

DETAILED ACTION

Claim Rejections - 35 USC § 112

The following is a quotation of the first paragraph of 35 U.S.C. 112:

The specification shall contain a written description of the invention, and of the manner and process of making and using it, in such full, clear, concise, and exact terms as to enable any person skilled in the art to which it pertains, or with which it is most nearly connected, to make and use the same and shall set forth the best mode contemplated by the inventor of carrying out his invention.

Claims 1, 10 and 16 are rejected under 35 U.S.C. 112, first paragraph, as failing to comply with the written description requirement. The claim(s) contains subject matter which was not described in the specification in such a way as to reasonably convey to one skilled in the relevant art that the inventor(s), at the time the application was filed, had possession of the claimed invention. The step of using milk surge data to subsequently effect one or more of a subsequent therapeutic, diagnostic or modified milk expression technique of the mother is not present in the specification as originally filed.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 1, 2, 4-10, and 12-18 are rejected under 35 U.S.C. 103(a) as being unpatentable over Alekseev et al (*Compression stimuli increase the efficacy of breast pump function*, European Journal of Obstetrics & Gynecology) as modified by Jaquith (US 4,030,356).

Alekseev et al disclose on page 133 a system for detecting a milk surge and using that data for modified milk expression techniques for a mother having a breast pump (1), a breast

shield (2), a collecting container (6), and a unit (volume, see page 133 column 1, lines 8-10) by which milk surge is detected over time.

Alekseev et al disclose the claimed invention except for the measuring means being an electromechanical balance with a bearing surface used to measure the weight of the milk. Jaquith teaches in column 2, lines 40-52 that it is known to use an electromechanical balance as a measuring means and weight (as opposed to volume) as a measuring unit. It would have been obvious to one having ordinary skill in the art at the time the invention was made to modify the system of Alekseev et al with an electromechanical balance as taught by Jaquith since such a modification would provide the system with a more efficient, accurate and automated way of measuring the weight of milk expressed.

Response to Arguments

Applicant's arguments filed 6/3/2008 have been fully considered but they are not persuasive. Applicant argues that the specification as filed provides basis for the limitation regarding the use of data to effect one or more of a subsequent therapeutic, diagnostic or modified milk expression technique of the mother. Although Applicant has shown specific examples of how the data can be used, these specific examples do not warrant the use of the broad limitation drawn to any therapeutic, diagnostic, or modified milk expression technique.

Applicant further argues the combination of Alekseev et al and Jaquith. Although Alekseev et al may not disclose the measurement of milk being based on weight, but rather on volume, Examiner points out that to one having ordinary skill in the art it would be obvious that a volume measurement could very easily be converted into a unit of weight of milk expressed (and vice versa). To this extent, it would be easier, and more accurate, to weigh the amount of milk expressed using a balance (as taught by Jaquith) than it would be to provide a volume measurement using a simple graduated container as taught by Alekseev et al. - but in the end,

both methods would provide the user with the same information about the quantity of milk expressed. For that reason it would be obvious to combine the two teachings.

Conclusion

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to CATHERINE N. WITCZAK whose telephone number is (571)272-7179. The examiner can normally be reached on Monday through Friday, 8-5 EST.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Kevin Sirmons can be reached on (571) 272-4965. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3767

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Catherine N Witezak/

Examiner, Art Unit 3767

/Kevin C. Simons/

Supervisory Patent Examiner, Art Unit 3767